
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.1uS / 0.0ppm
CYCLE NUMBER 008282
11:14 31/01/2025
WRAPPED
135 C for 10 min.
CYCLE START 0:00
CONDITIONING 2:27
PRESSURIZING 4:25
STERILIZING 5:17
136.4 C 221kPa 5:17
136.4 C 224kPa 5:47
136.3 C 222kPa 6:17
136.4 C 223kPa 6:47
136.2 C 223kPa 7:17
136.4 C 224kPa 7:47
136.4 C 221kPa 8:17
136.3 C 222kPa 8:47
136.4 C 222kPa 9:17
136.1 C 221kPa 9:47
136.4 C 225kPa 10:17
136.3 C 224kPa 10:47
136.4 C 222kPa 11:17
CYCLE ABORTED 11:27
STOP BUTTON PRESSED
NOT STERILE
END TIME 11:27

Digital Signature #
15E624069D8FCB12
