
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 002155
9:09 28/07/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:20
PRESSURIZING 2:22
STERILIZING 3:17
136.4 C 221kPa 3:17
136.3 C 223kPa 3:47
136.2 C 222kPa 4:17
136.3 C 223kPa 4:47
136.3 C 222kPa 5:17
136.4 C 222kPa 5:47
136.4 C 220kPa 6:17
136.4 C 223kPa 6:47
136.4 C 223kPa 6:47
VENTING 6:47
AIR DRYING 7:39
CYCLE COMPLETE 7:41

Digital Signature #
6A5B1CA04C317F14
