
Midmark M11 - v1.0.0
Total Cycles: 2321

Name: RITTER 1
Sterilizer ID: V2178198
09/13/2022 02:36 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 159.1 F 0.0
2:00 159.7 F 2.8
4:00 168.8 F 6.6
6:00 190.5 F 11.6
8:00 217.0 F 5.6
10:00 232.1 F 10.6
12:00 235.5 F 7.1
14:00 241.6 F 12.9
16:00 240.2 F 11.2
18:00 253.3 F 17.7
20:00 264.9 F 24.5

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 29.0
1:00 272.7 F 29.6
2:00 271.7 F 29.1
3:00 272.7 F 29.6
4:00 271.7 F 28.9

Min 271.6 F 28.9
Max 272.9 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 212.4 F 0.0
5:00 199.2 F 0.0
10:00 212.3 F 0.0
15:00 219.8 F 0.0
20:00 221.9 F 0.0
25:00 212.3 F 0.0
30:00 207.9 F 0.0
Drying Complete

FILLING: 1:40
HEATING: 21:52
STERILIZING: 4:00
VENTING: 3:12
DRYING: 30:00
TOTAL CYCLE: 01:00:44
CYCLE COMPLETE

09/13/2022 03:36 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____