
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 001025
16:04 06/04/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:55
PRESSURIZING 2:02
STERILIZING 3:08
136.0 C 223kPa 3:08
136.3 C 225kPa 3:38
136.4 C 224kPa 4:08
136.4 C 225kPa 4:38
136.4 C 226kPa 5:08
136.3 C 224kPa 5:38
136.3 C 225kPa 6:08
136.2 C 224kPa 6:38
136.2 C 224kPa 6:38
VENTING 6:38
AIR DRYING 7:32
CYCLE COMPLETE 19:34

Digital Signature #
367D9D8DF4EB04B2
