
Midmark M11 - v1.0.5
Total Cycles: 2052

Name: RITTER2
Sterilizer ID: V2313228
02/14/2023 06:30 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 184.9 F 0.0
2:00 180.7 F 2.9
4:00 193.4 F 7.1
6:00 212.9 F 12.4
8:00 224.3 F 7.7
10:00 237.5 F 13.3
12:00 239.6 F 11.3
14:00 239.2 F 10.4
16:00 253.0 F 17.2
18:00 264.8 F 24.3

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.9
1:00 272.5 F 29.3
2:00 271.8 F 29.3
3:00 272.3 F 29.2
4:00 271.6 F 29.1

Min 271.5 F 28.7
Max 272.8 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 202.9 F 0.0
5:00 216.9 F 0.0
10:00 226.0 F 0.0
15:00 229.2 F 0.0
20:00 230.0 F 0.0
25:00 228.2 F 0.0
30:00 225.8 F 0.0
Drying Complete

FILLING: 1:44
HEATING: 19:50
STERILIZING: 4:00
VENTING: 3:16
DRYING: 30:00
TOTAL CYCLE: 00:58:51
CYCLE COMPLETE

02/14/2023 07:28 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____