
Midmark M11 - v1.0.0
Total Cycles: 1323

Name: RITTER 1
Sterilizer ID: V2178198
10/29/2021 01:41 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 149.8 F 0.0
2:00 155.0 F 2.4
4:00 182.2 F 6.2
6:00 203.4 F 11.4
8:00 224.7 F 6.4
10:00 237.4 F 12.0
12:00 236.6 F 9.1
14:00 237.5 F 9.3
16:00 248.3 F 14.2
18:00 260.8 F 21.2
20:00 272.3 F 28.8

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.3
1:00 271.7 F 28.3
2:00 272.3 F 28.6
3:00 272.6 F 28.9
4:00 271.6 F 28.6

Min 271.2 F 28.2
Max 272.7 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 203.6 F 0.0
5:00 213.3 F 0.0
10:00 195.8 F 0.0
15:00 193.2 F 0.0
20:00 189.4 F 0.0
25:00 186.2 F 0.0
30:00 183.6 F 0.0
Drying Complete

FILLING: 1:38
HEATING: 20:24
STERILIZING: 4:00
VENTING: 3:09
DRYING: 30:00
TOTAL CYCLE: 00:59:10
CYCLE COMPLETE

10/29/2021 03:25 PM
ACCEPTED: X
REJECTED:
Operator: U02