
Midmark M11 - v1.0.0
Total Cycles: 2596

Name: RITTER 1
Sterilizer ID: V2178198
11/22/2022 01:46 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 217.1 F 0.0
2:00 200.9 F 4.8
4:00 226.8 F 10.7
6:00 228.7 F 6.3
8:00 240.1 F 12.0
10:00 238.0 F 9.7
12:00 237.5 F 8.5
14:00 251.1 F 15.6
16:00 263.7 F 22.8

STERILIZING
mm:ss Degrees PSI
0:00 271.8 F 28.1
1:00 272.3 F 28.7
2:00 271.2 F 28.1
3:00 272.2 F 28.6
4:00 271.2 F 28.1

Min 271.1 F 27.9
Max 272.5 F 28.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 208.5 F 0.0
5:00 232.2 F 0.0
10:00 225.7 F 0.0
15:00 222.7 F 0.0
20:00 232.2 F 0.0
25:00 227.3 F 0.0
30:00 231.7 F 0.0
Drying Complete

FILLING: 2:46
HEATING: 17:50
STERILIZING: 4:00
VENTING: 3:15
DRYING: 30:00
TOTAL CYCLE: 00:57:50
CYCLE COMPLETE

11/22/2022 02:43 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____