
Midmark M11 - v1.0.0
Total Cycles: 2201

Name: RITTER 1
Sterilizer ID: V2178198
08/04/2022 12:48 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 138.5 F 0.0
2:00 138.7 F 1.1
4:00 161.3 F 4.3
6:00 182.2 F 8.3
8:00 203.7 F 13.0
10:00 217.0 F 7.0
12:00 230.0 F 11.7
14:00 230.3 F 7.7
16:00 242.4 F 13.2
18:00 241.0 F 11.2
20:00 253.4 F 17.4
22:00 264.3 F 23.8
24:00 271.5 F 28.6

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.8
1:00 272.6 F 29.4
2:00 271.6 F 28.6
3:00 272.3 F 29.5
4:00 272.1 F 29.1

Min 271.3 F 28.6
Max 272.6 F 29.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 187.0 F 0.0
5:00 195.3 F 0.0
10:00 213.4 F 0.0
15:00 219.9 F 0.0
20:00 222.7 F 0.0
25:00 223.0 F 0.0
30:00 220.5 F 0.0
Drying Complete

FILLING: 0:56
HEATING: 24:05
STERILIZING: 4:00
VENTING: 3:12
DRYING: 30:00
TOTAL CYCLE: 01:02:13
CYCLE COMPLETE

08/04/2022 01:50 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____