

-----  
Midmark M11 - v1.0.5  
Total Cycles: 1395

Name: RITTER2  
Sterilizer ID: V2313228  
08/16/2022 08:03 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 158.9 F 0.0  
2:00 158.9 F 0.0  
4:00 165.8 F 1.5  
6:00 187.5 F 6.3  
8:00 209.4 F 11.2  
10:00 220.6 F 6.1  
12:00 234.2 F 11.4  
14:00 234.3 F 9.0  
16:00 239.4 F 10.3  
18:00 247.3 F 14.4  
20:00 259.5 F 21.2  
22:00 270.6 F 28.4

STERILIZING  
mm:ss Degrees PSI  
0:00 271.9 F 29.0  
1:00 272.8 F 29.6  
2:00 271.9 F 29.1  
3:00 272.9 F 29.7  
4:00 272.1 F 29.0

Min 271.9 F 28.9  
Max 272.9 F 29.9

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 206.5 F 0.0  
5:00 200.2 F 0.0  
10:00 207.1 F 0.0  
15:00 211.0 F 0.0  
20:00 211.5 F 0.0  
25:00 211.3 F 0.0  
30:00 213.0 F 0.0  
Drying Complete

FILLING: 1:23  
HEATING: 22:47  
STERILIZING: 4:00  
VENTING: 3:10  
DRYING: 30:00  
TOTAL CYCLE: 01:01:20  
CYCLE COMPLETE

08/16/2022 09:04 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_