
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 005081
9:48 07/02/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:28
PRESSURIZING 3:28
STERILIZING 4:28
136.4 C 221kPa 4:28
136.4 C 224kPa 4:58
136.4 C 219kPa 5:28
136.4 C 223kPa 5:58
136.2 C 219kPa 6:28
136.2 C 220kPa 6:58
136.4 C 222kPa 7:28
135.8 C 218kPa 7:58
135.8 C 218kPa 7:58
VENTING 7:58
AIR DRYING 9:11
CYCLE COMPLETE 11:36

Digital Signature #
C2CBC38EE48DCCD9
