
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 000138
18:52 08/09/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:56
PRESSURIZING 3:00
STERILIZING 4:10
136.0 C 222kPa 4:10
136.3 C 224kPa 4:40
136.3 C 225kPa 5:10
136.2 C 223kPa 5:40
136.4 C 226kPa 6:10
136.4 C 226kPa 6:40
136.4 C 223kPa 7:10
136.2 C 225kPa 7:40
136.2 C 225kPa 7:40
VENTING 7:40
AIR DRYING 8:34
CYCLE COMPLETE 10:39

Digital Signature #
E02577627FE0A0EE
