
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 001254
12:55 14/09/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 4:10
PRESSURIZING 4:11
STERILIZING 5:23
136.0 C 220kPa 5:23
136.4 C 224kPa 5:53
136.3 C 223kPa 6:23
136.3 C 223kPa 6:53
136.1 C 221kPa 7:23
136.3 C 224kPa 7:53
136.4 C 222kPa 8:23
136.4 C 224kPa 8:53
136.4 C 224kPa 8:53
VENTING 8:53
AIR DRYING 9:52
CYCLE COMPLETE 24:52

Digital Signature #
1F669EB19B68E489
