

-----  
Midmark M11 - v1.0.5  
Total Cycles: 2845

Name: RITTER2  
Sterilizer ID: V2313228  
09/07/2023 01:07 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 179.9 F 0.0  
2:00 180.3 F 3.3  
4:00 193.1 F 8.0  
6:00 210.4 F 12.9  
8:00 224.0 F 8.1  
10:00 236.7 F 13.5  
12:00 239.2 F 11.5  
14:00 239.0 F 10.6  
16:00 251.8 F 17.0  
18:00 263.1 F 23.6  
20:00 272.1 F 29.3

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 29.3  
1:00 272.0 F 29.2  
2:00 272.2 F 29.6  
3:00 272.2 F 29.9  
4:00 271.1 F 29.3

Min 271.1 F 29.1  
Max 272.5 F 29.9

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 197.1 F 0.0  
5:00 218.1 F 0.0  
10:00 218.7 F 0.0  
15:00 216.1 F 0.0  
20:00 211.9 F 0.0  
25:00 216.1 F 0.0  
30:00 215.8 F 0.0  
Drying Complete

FILLING: 1:58  
HEATING: 20:08  
STERILIZING: 4:00  
VENTING: 3:38  
DRYING: 30:00  
TOTAL CYCLE: 00:59:44  
CYCLE COMPLETE

09/07/2023 02:06 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_