

-----  
Midmark M11 - v1.0.0  
Total Cycles: 2236

Name: RITTER 1  
Sterilizer ID: V2178198  
08/24/2022 07:29 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 71.8 F 0.0  
2:00 73.7 F 0.0  
4:00 96.0 F 0.0  
6:00 137.6 F 0.0  
8:00 148.4 F 0.2  
10:00 186.3 F 0.5  
12:00 203.8 F 2.1  
14:00 217.9 F 5.2  
16:00 228.6 F 8.2  
18:00 237.9 F 11.7  
20:00 230.1 F 7.0  
22:00 241.5 F 12.1  
24:00 237.7 F 9.6  
26:00 238.7 F 9.1  
28:00 248.7 F 14.4  
30:00 260.0 F 20.7  
32:00 270.4 F 27.3

STERILIZING  
mm:ss Degrees PSI  
0:00 272.0 F 28.1  
1:00 272.4 F 28.5  
2:00 272.6 F 28.8  
3:00 272.0 F 28.5  
4:00 271.7 F 28.0  
  
Min 271.5 F 27.9  
Max 272.7 F 28.9

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 203.7 F 0.0  
5:00 217.1 F 0.0  
10:00 238.8 F 0.0  
15:00 228.7 F 0.0  
20:00 216.1 F 0.0  
25:00 208.7 F 0.0  
30:00 207.0 F 0.0  
  
Drying Complete

FILLING: 0:45  
HEATING: 32:42  
STERILIZING: 4:00  
VENTING: 3:14  
DRYING: 30:00  
TOTAL CYCLE: 01:10:41  
CYCLE COMPLETE

08/24/2022 08:39 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_