
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R750
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 000596
10:37 07/10/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 4:00
PRESSURIZING 4:01
STERILIZING 5:08
136.4 C 222kPa 5:08
136.0 C 220kPa 5:38
136.2 C 222kPa 6:08
136.4 C 222kPa 6:38
136.4 C 223kPa 7:08
136.4 C 224kPa 7:38
136.3 C 222kPa 8:08
136.3 C 222kPa 8:38
136.3 C 222kPa 8:38
VENTING 8:38
AIR DRYING 9:32
CYCLE COMPLETE 9:34

Digital Signature #
1AA3C243D117F6B0
