
Midmark M11 - v1.0.5
Total Cycles: 1499

Name: RITTER2
Sterilizer ID: V2313228
09/19/2022 09:06 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 140.6 F 0.0
2:00 140.6 F 1.0
4:00 153.5 F 3.7
6:00 174.8 F 7.7
8:00 197.4 F 12.5
10:00 216.6 F 6.7
12:00 230.1 F 11.7
14:00 231.5 F 8.5
16:00 239.0 F 11.1
18:00 244.3 F 13.1
20:00 256.3 F 19.6
22:00 267.6 F 26.7

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 29.3
1:00 272.4 F 30.2
2:00 272.2 F 29.6
3:00 271.1 F 29.2
4:00 271.7 F 29.9

Min 271.1 F 29.2
Max 272.6 F 30.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 192.9 F 0.0
5:00 215.4 F 0.0
10:00 232.1 F 0.0
15:00 231.8 F 0.0
20:00 232.8 F 0.0
25:00 228.6 F 0.0
30:00 230.1 F 0.0
Drying Complete

FILLING: 0:59
HEATING: 23:25
STERILIZING: 4:00
VENTING: 3:10
DRYING: 30:00
TOTAL CYCLE: 01:01:34
CYCLE COMPLETE

09/19/2022 10:07 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____