
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.1uS / 0.0ppm
CYCLE NUMBER 001540
10:25 24/10/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:48
PRESSURIZING 2:52
STERILIZING 3:54
136.0 C 222kPa 3:54
136.0 C 223kPa 4:24
136.1 C 223kPa 4:54
136.3 C 225kPa 5:24
136.4 C 226kPa 5:54
136.4 C 226kPa 6:24
136.3 C 224kPa 6:54
136.4 C 225kPa 7:24
136.4 C 225kPa 7:24
VENTING 7:24
AIR DRYING 8:14
CYCLE COMPLETE 9:35

Digital Signature #
A4F65BF7869312D4
