

-----  
Midmark M11 - v1.0.5  
Total Cycles: 48

Name: RITTER2  
Sterilizer ID: V2313228  
05/27/2021 09:10 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 147.7 F 0.0  
2:00 148.2 F 1.9  
4:00 153.1 F 4.6  
6:00 173.5 F 8.3  
8:00 195.8 F 12.9  
10:00 213.1 F 6.4  
12:00 225.6 F 11.1  
14:00 230.3 F 8.0  
16:00 242.5 F 13.5  
18:00 243.1 F 12.5  
20:00 254.9 F 18.6  
22:00 265.2 F 25.1

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 29.2  
1:00 272.4 F 30.1  
2:00 272.3 F 29.5  
3:00 271.5 F 29.1  
4:00 272.2 F 30.0

Min 271.5 F 29.1  
Max 272.9 F 30.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 184.7 F 0.0  
5:00 202.0 F 0.0  
10:00 196.7 F 0.0  
15:00 199.4 F 0.0  
20:00 201.5 F 0.0  
25:00 204.4 F 0.0  
30:00 198.2 F 0.0  
Drying Complete

FILLING: 1:22  
HEATING: 23:56  
STERILIZING: 4:00  
VENTING: 3:23  
DRYING: 30:00  
TOTAL CYCLE: 01:02:40  
CYCLE COMPLETE

05/27/2021 10:12 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_