
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 004989
16:52 23/01/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:49
PRESSURIZING 1:53
STERILIZING 2:54
136.4 C 223kPa 2:54
136.0 C 220kPa 3:24
136.4 C 223kPa 3:54
136.4 C 222kPa 4:24
136.4 C 224kPa 4:54
136.4 C 225kPa 5:24
136.3 C 224kPa 5:54
136.3 C 222kPa 6:24
136.3 C 222kPa 6:24
VENTING 6:24
AIR DRYING 7:15
CYCLE COMPLETE 10:29

Digital Signature #
79F2B84A311A7748
