
Midmark M11 - v1.0.0
Total Cycles: 5351

Name: RITTER 1
Sterilizer ID: V2178198
02/20/2025 05:10 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 140.1 F 0.0
2:00 142.4 F 1.4
4:00 159.2 F 3.0
6:00 189.8 F 5.0
8:00 208.7 F 7.4
10:00 221.2 F 10.1
12:00 233.8 F 13.0
14:00 236.1 F 9.3
16:00 239.7 F 9.8
18:00 246.3 F 13.3
20:00 244.4 F 12.3
22:00 257.1 F 18.9
24:00 268.4 F 25.9

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.0
1:00 271.7 F 28.6
2:00 272.7 F 29.0
3:00 271.8 F 28.4
4:00 272.0 F 28.8

Min 270.9 F 27.9
Max 272.9 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 186.3 F 0.0
5:00 214.8 F 0.0
10:00 191.6 F 0.0
15:00 191.4 F 0.0
20:00 190.0 F 0.0
25:00 188.5 F 0.0
30:00 190.4 F 0.0
Drying Complete

FILLING: 1:02
HEATING: 25:09
STERILIZING: 4:00
VENTING: 3:42
DRYING: 30:00
TOTAL CYCLE: 01:03:53
CYCLE COMPLETE

02/20/2025 06:13 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____