
Midmark M11 - v1.0.5
Total Cycles: 201

Name: RITTER2
Sterilizer ID: V2313228
07/13/2021 12:15 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 109.6 F 0.0
2:00 113.1 F 1.2
4:00 132.6 F 3.9
6:00 161.8 F 7.7
8:00 186.0 F 12.1
10:00 209.9 F 5.4
12:00 219.7 F 9.9
14:00 233.4 F 8.9
16:00 238.0 F 12.0
18:00 239.2 F 10.2
20:00 251.3 F 16.3
22:00 261.2 F 22.5
24:00 271.6 F 29.2

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 29.5
1:00 271.8 F 29.7
2:00 271.8 F 29.8
3:00 271.6 F 29.8
4:00 271.6 F 29.6

Min 271.6 F 29.4
Max 272.9 F 30.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 182.3 F 0.0
5:00 231.4 F 0.0
10:00 196.5 F 0.0
15:00 211.1 F 0.0
20:00 214.2 F 0.0
25:00 233.5 F 0.0
30:00 225.0 F 0.0
Drying Complete

FILLING: 1:03
HEATING: 25:12
STERILIZING: 4:00
VENTING: 3:08
DRYING: 30:00
TOTAL CYCLE: 01:03:23
CYCLE COMPLETE

07/13/2021 01:19 PM
ACCEPTED: X
REJECTED:
Operator: U01