
Midmark M11 - v1.0.5
Total Cycles: 1227

Name: RITTER2
Sterilizer ID: V2313228
06/27/2022 10:55 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 201.9 F 0.0
2:00 197.2 F 4.1
4:00 215.8 F 9.4
6:00 224.1 F 8.0
8:00 232.9 F 10.9
10:00 233.4 F 8.6
12:00 239.0 F 10.3
14:00 247.5 F 14.6
16:00 260.1 F 21.5
18:00 271.3 F 29.0

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 29.2
1:00 272.3 F 29.5
2:00 271.7 F 29.5
3:00 272.3 F 29.5
4:00 271.4 F 29.3

Min 271.3 F 29.0
Max 272.5 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 190.2 F 0.0
5:00 221.2 F 0.0
10:00 241.3 F 0.0
15:00 221.4 F 0.0
20:00 219.7 F 0.0
25:00 220.6 F 0.0
30:00 219.7 F 0.0
Drying Complete

FILLING: 1:57
HEATING: 18:39
STERILIZING: 4:00
VENTING: 3:11
DRYING: 30:00
TOTAL CYCLE: 00:57:47
CYCLE COMPLETE

06/27/2022 11:52 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____