
Midmark M11 - v1.0.0
Total Cycles: 2142

Name: RITTER 1
Sterilizer ID: V2178198
07/18/2022 09:12 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 165.8 F 0.0
2:00 160.7 F 0.0
4:00 161.9 F 2.9
6:00 193.0 F 6.8
8:00 204.7 F 11.6
10:00 222.2 F 5.2
12:00 233.5 F 9.7
14:00 238.2 F 9.7
16:00 241.5 F 11.5
18:00 238.1 F 9.4
20:00 251.3 F 15.9
22:00 262.7 F 22.4
24:00 272.8 F 28.9

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.2
1:00 272.7 F 29.1
2:00 272.0 F 28.4
3:00 271.4 F 28.5
4:00 272.4 F 29.1

Min 271.2 F 28.1
Max 272.8 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 198.3 F 0.0
5:00 234.6 F 0.0
10:00 231.3 F 0.0
15:00 231.7 F 0.0
20:00 227.3 F 0.0
25:00 225.9 F 0.0
30:00 233.1 F 0.0
Drying Complete

FILLING: 0:46
HEATING: 24:22
STERILIZING: 4:00
VENTING: 3:11
DRYING: 30:00
TOTAL CYCLE: 01:02:19
CYCLE COMPLETE

07/18/2022 10:14 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____