
Midmark M11 - v1.0.0
Total Cycles: 1696

Name: RITTER 1
Sterilizer ID: V2178198
03/02/2022 04:45 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 187.9 F 0.0
2:00 184.9 F 3.6
4:00 195.9 F 8.2
6:00 222.1 F 13.4
8:00 225.6 F 8.3
10:00 239.0 F 11.9
12:00 239.7 F 11.3
14:00 238.1 F 9.6
16:00 252.6 F 16.5
18:00 264.0 F 23.3

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.2
1:00 271.8 F 28.7
2:00 272.9 F 29.5
3:00 272.4 F 29.0
4:00 271.8 F 28.6

Min 270.9 F 28.1
Max 273.1 F 29.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 211.0 F 0.0
5:00 223.3 F 0.0
10:00 228.7 F 0.0
15:00 231.6 F 0.0
20:00 230.1 F 0.0
25:00 225.4 F 0.0
30:00 225.6 F 0.0
Drying Complete

FILLING: 1:50
HEATING: 19:56
STERILIZING: 4:00
VENTING: 3:08
DRYING: 30:00
TOTAL CYCLE: 00:58:53
CYCLE COMPLETE

03/02/2022 05:43 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____