

-----  
Midmark M11 - v1.0.0  
Total Cycles: 5065

Name: RITTER 1  
Sterilizer ID: V2178198  
11/07/2024 09:44 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 69.5 F 0.0  
2:00 73.5 F 0.0  
4:00 86.2 F 0.0  
6:00 120.3 F 0.1  
8:00 148.5 F 0.3  
10:00 174.6 F 0.9  
12:00 192.8 F 3.0  
14:00 200.6 F 5.8  
16:00 210.6 F 8.6  
18:00 221.1 F 11.8  
20:00 222.0 F 5.2  
22:00 232.7 F 9.7  
24:00 239.4 F 10.8  
26:00 241.6 F 11.3  
28:00 237.7 F 9.3  
30:00 250.4 F 15.2  
32:00 260.8 F 21.0  
34:00 269.9 F 27.1

STERILIZING  
mm:ss Degrees PSI  
0:00 271.6 F 28.1  
1:00 272.2 F 28.6  
2:00 272.7 F 28.9  
3:00 272.3 F 28.5  
4:00 271.8 F 28.2  
  
Min 271.5 F 28.0  
Max 272.7 F 29.0

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 177.5 F 0.0  
5:00 205.3 F 0.0  
10:00 205.6 F 0.0  
15:00 219.7 F 0.0  
20:00 235.9 F 0.0  
25:00 233.6 F 0.0  
30:00 223.3 F 0.0  
Drying Complete

FILLING: 0:41  
HEATING: 35:07  
STERILIZING: 4:00  
VENTING: 3:48  
DRYING: 30:00  
TOTAL CYCLE: 01:13:37  
CYCLE COMPLETE

11/07/2024 10:57 AM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_