
Midmark M11 - v1.0.5
Total Cycles: 187

Name: RITTER2
Sterilizer ID: V2313228
07/08/2021 04:22 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 162.8 F 0.0
2:00 176.5 F 4.1
4:00 202.9 F 10.0
6:00 219.5 F 5.7
8:00 235.8 F 12.7
10:00 238.9 F 11.4
12:00 240.8 F 11.0
14:00 255.5 F 18.5
16:00 267.8 F 26.2

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.7
1:00 272.4 F 29.4
2:00 271.2 F 28.8
3:00 272.4 F 29.5
4:00 271.5 F 28.7

Min 271.2 F 28.6
Max 272.7 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 175.5 F 0.0
5:00 231.3 F 0.0
10:00 203.5 F 0.0
15:00 199.8 F 0.0
20:00 189.2 F 0.0
Dry Cycle Aborted

FILLING: 3:38
HEATING: 17:15
STERILIZING: 4:00
VENTING: 2:56
DRYING: 24:10
TOTAL CYCLE: 00:51:58

07/08/2021 05:14 PM
ACCEPTED: X
REJECTED:
Operator: U02