
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 000478
14:11 26/02/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:46
PRESSURIZING 3:50
STERILIZING 5:05
136.0 C 228kPa 5:05
136.3 C 232kPa 5:35
136.4 C 233kPa 6:05
136.2 C 231kPa 6:35
136.4 C 233kPa 7:05
136.4 C 232kPa 7:35
136.4 C 230kPa 8:05
136.1 C 230kPa 8:35
136.1 C 230kPa 8:35
VENTING 8:35
AIR DRYING 9:29
CASSETTE REMOVED
CYCLE COMPLETE 14:45

Digital Signature #
1EFD0171A4AE657E
