

-----  
Midmark M11 - v1.0.5  
Total Cycles: 3183

Name: RITTER2  
Sterilizer ID: V2313228  
12/05/2023 09:40 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 69.3 F 0.0  
2:00 69.6 F 0.0  
4:00 72.8 F 0.2  
6:00 97.7 F 1.2  
8:00 134.1 F 3.6  
10:00 162.6 F 6.2  
12:00 181.8 F 9.3  
14:00 199.1 F 12.6  
16:00 210.3 F 5.7  
18:00 219.8 F 9.4  
20:00 231.7 F 12.6  
22:00 233.5 F 9.8  
24:00 237.4 F 9.6  
26:00 244.2 F 13.1  
28:00 255.1 F 18.8  
30:00 265.1 F 24.9

STERILIZING  
mm:ss Degrees PSI  
0:00 271.2 F 29.3  
1:00 272.5 F 30.0  
2:00 271.5 F 29.2  
3:00 272.2 F 30.1  
4:00 272.2 F 29.6  
  
Min 271.2 F 29.1  
Max 272.8 F 30.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 211.0 F 0.0  
5:00 214.3 F 0.0  
10:00 193.9 F 0.0  
15:00 188.5 F 0.0  
20:00 184.1 F 0.0  
25:00 181.3 F 0.0  
30:00 179.1 F 0.0  
Drying Complete

FILLING: 0:48  
HEATING: 31:59  
STERILIZING: 4:00  
VENTING: 3:24  
DRYING: 30:00  
TOTAL CYCLE: 01:10:12  
CYCLE COMPLETE

12/05/2023 10:50 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_