

-----  
Midmark M11 - v1.0.5  
Total Cycles: 190

Name: RITTER2  
Sterilizer ID: V2313228  
07/12/2021 07:36 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 184.0 F 0.0  
2:00 182.0 F 2.0  
4:00 187.9 F 6.1  
6:00 207.1 F 11.6  
8:00 221.2 F 6.8  
10:00 234.6 F 12.1  
12:00 236.6 F 10.0  
14:00 237.3 F 9.0  
16:00 250.3 F 15.8  
18:00 262.1 F 22.5  
20:00 272.8 F 29.5

STERILIZING  
mm:ss Degrees PSI  
0:00 272.0 F 28.8  
1:00 272.6 F 29.2  
2:00 272.7 F 29.6  
3:00 271.8 F 29.0  
4:00 272.3 F 29.0

Min 271.6 F 28.6  
Max 272.9 F 29.6

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 198.9 F 0.0  
5:00 208.3 F 0.0  
10:00 220.0 F 0.0  
15:00 224.5 F 0.0  
20:00 221.7 F 0.0  
25:00 219.6 F 0.0  
30:00 216.7 F 0.0  
Drying Complete

FILLING: 2:20  
HEATING: 20:12  
STERILIZING: 4:00  
VENTING: 3:04  
DRYING: 30:00  
TOTAL CYCLE: 00:59:36  
CYCLE COMPLETE

07/12/2021 08:35 AM  
ACCEPTED: X  
REJECTED:  
Operator: U01