
Midmark M11 - v1.0.5
Total Cycles: 3492

Name: RITTER2
Sterilizer ID: V2313228
04/03/2024 08:45 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 68.8 F 0.0
2:00 69.8 F 0.0
4:00 74.1 F 0.0
6:00 109.3 F 0.1
8:00 137.7 F 1.0
10:00 168.0 F 3.8
12:00 183.7 F 7.3
14:00 203.6 F 11.0
16:00 217.3 F 6.2
18:00 224.3 F 8.9
20:00 234.5 F 13.1
22:00 235.6 F 10.0
24:00 235.9 F 8.9
26:00 246.9 F 14.3
28:00 258.3 F 20.5
30:00 268.5 F 27.1

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 29.2
1:00 272.2 F 29.5
2:00 271.9 F 29.9
3:00 271.7 F 29.2
4:00 272.3 F 29.9

Min 271.3 F 29.1
Max 272.4 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 184.1 F 0.0
5:00 218.8 F 0.0
10:00 224.2 F 0.0
15:00 201.1 F 0.0
20:00 190.2 F 0.0
25:00 187.5 F 0.0
30:00 192.1 F 0.0
Drying Complete

FILLING: 0:51
HEATING: 31:13
STERILIZING: 4:00
VENTING: 3:18
DRYING: 30:00
TOTAL CYCLE: 01:09:23
CYCLE COMPLETE

04/03/2024 09:54 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____