
Midmark M11 - v1.0.0
Total Cycles: 787

Name: RITTER 1
Sterilizer ID: V2178198
04/15/2021 11:49 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 119.2 F 0.0
2:00 120.2 F 0.7
4:00 135.4 F 2.4
6:00 167.2 F 5.9
8:00 200.6 F 10.7
10:00 215.8 F 4.6
12:00 226.4 F 9.9
14:00 234.0 F 8.1
16:00 241.5 F 12.4
18:00 240.3 F 10.7
20:00 253.8 F 17.4
22:00 264.8 F 24.5

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 28.5
1:00 272.7 F 29.2
2:00 271.1 F 28.6
3:00 272.6 F 29.2
4:00 271.2 F 28.4

Min 271.1 F 28.4
Max 272.9 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 194.3 F 0.0
5:00 232.0 F 0.0
10:00 219.8 F 0.0
15:00 218.2 F 0.0
20:00 238.0 F 0.0
25:00 236.4 F 0.0
30:00 226.3 F 0.0
Drying Complete

FILLING: 1:02
HEATING: 23:42
STERILIZING: 4:00
VENTING: 3:01
DRYING: 30:00
TOTAL CYCLE: 01:01:44
CYCLE COMPLETE

04/15/2021 12:50 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____