
Midmark M11 - v1.0.0
Total Cycles: 2816

Name: RITTER 1
Sterilizer ID: V2178198
01/25/2023 08:26 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 78.8 F 0.0
2:00 80.5 F 0.0
4:00 95.0 F 0.0
6:00 130.6 F 0.0
8:00 158.4 F 0.0
10:00 185.7 F 2.4
12:00 201.8 F 5.9
14:00 213.7 F 9.2
16:00 223.9 F 12.6
18:00 226.5 F 6.5
20:00 235.1 F 10.8
22:00 235.7 F 7.4
24:00 244.2 F 12.7
26:00 241.5 F 10.8
28:00 253.6 F 17.0
30:00 264.3 F 23.2
32:00 271.7 F 27.9

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.1
1:00 272.5 F 28.7
2:00 271.6 F 27.9
3:00 272.5 F 28.8
4:00 272.3 F 28.4

Min 271.4 F 27.9
Max 272.7 F 28.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 209.4 F 0.0
5:00 210.7 F 0.0
10:00 220.2 F 0.0
15:00 224.0 F 0.0
20:00 223.8 F 0.0
25:00 225.3 F 0.0
30:00 224.8 F 0.0

Drying Complete

FILLING: 0:46
HEATING: 32:08
STERILIZING: 4:00
VENTING: 3:13
DRYING: 30:00
TOTAL CYCLE: 01:10:06
CYCLE COMPLETE

01/25/2023 09:36 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____