
Midmark M11 - v1.0.0
Total Cycles: 1332

Name: RITTER 1
Sterilizer ID: V2178198
11/02/2021 12:07 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 204.8 F 0.0
2:00 188.0 F 3.2
4:00 206.4 F 8.2
6:00 235.1 F 9.1
8:00 231.1 F 10.4
10:00 232.1 F 7.2
12:00 240.9 F 10.7
14:00 247.1 F 13.4
16:00 260.0 F 20.6
18:00 271.7 F 28.3

STERILIZING
mm:ss Degrees PSI
0:00 272.2 F 28.6
1:00 271.8 F 28.5
2:00 271.7 F 28.8
3:00 271.7 F 28.8
4:00 271.6 F 28.8

Min 271.5 F 28.5
Max 272.8 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 202.8 F 0.0
5:00 229.8 F 0.0
10:00 225.2 F 0.0
15:00 238.3 F 0.0
20:00 228.2 F 0.0
25:00 231.4 F 0.0
30:00 227.0 F 0.0

Drying Complete

FILLING: 1:58
HEATING: 19:06
STERILIZING: 4:00
VENTING: 3:08
DRYING: 30:00
TOTAL CYCLE: 00:58:11
CYCLE COMPLETE

11/03/2021 06:33 AM
ACCEPTED: X
REJECTED:
Operator: U01