

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.4uS / 0.2ppm  
CYCLE NUMBER 000899  
15:39 02/02/2022  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 1:17  
PRESSURIZING 1:29  
STERILIZING 2:34  
136.4 C 231kPa 2:34  
136.3 C 232kPa 3:04  
136.3 C 232kPa 3:34  
136.4 C 233kPa 4:04  
136.2 C 230kPa 4:34  
136.4 C 230kPa 5:04  
136.4 C 230kPa 5:34  
136.4 C 233kPa 6:04  
136.4 C 233kPa 6:04  
VENTING 6:04  
AIR DRYING 6:56  
CYCLE COMPLETE 15:50

-----  
Digital Signature #  
FF323D82D08D01E3  
-----