
Midmark M11 - v1.0.5
Total Cycles: 2105

Name: RITTER2
Sterilizer ID: V2313228
02/28/2023 06:13 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 180.4 F 0.0
2:00 183.1 F 3.6
4:00 202.1 F 9.0
6:00 220.6 F 7.5
8:00 233.2 F 10.4
10:00 233.7 F 8.3
12:00 239.4 F 10.2
14:00 249.3 F 15.3
16:00 262.4 F 22.8

STERILIZING
mm:ss Degrees PSI
0:00 272.1 F 28.9
1:00 272.7 F 29.5
2:00 271.8 F 29.2
3:00 272.4 F 29.3
4:00 271.9 F 29.3

Min 271.6 F 28.8
Max 273.0 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 192.9 F 0.0
5:00 212.2 F 0.0
10:00 217.4 F 0.0
15:00 223.2 F 0.0
20:00 224.1 F 0.0
25:00 228.8 F 0.0
30:00 224.8 F 0.0

Drying Complete

FILLING: 2:23
HEATING: 17:59
STERILIZING: 4:00
VENTING: 3:16
DRYING: 30:00
TOTAL CYCLE: 00:57:39
CYCLE COMPLETE

02/28/2023 07:10 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____