
Midmark M11 - v1.0.0
Total Cycles: 3239

Name: RITTER 1
Sterilizer ID: V2178198
05/16/2023 09:29 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 201.8 F 0.0
2:00 194.4 F 0.7
4:00 217.2 F 6.3
6:00 229.5 F 11.9
8:00 231.9 F 8.4
10:00 241.0 F 11.0
12:00 244.0 F 12.8
14:00 243.5 F 12.1
16:00 257.2 F 19.3
18:00 269.0 F 26.7

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.4
1:00 272.5 F 29.0
2:00 271.5 F 28.4
3:00 272.7 F 29.2
4:00 271.9 F 28.5

Min 271.5 F 28.3
Max 272.8 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 210.6 F 0.0
5:00 217.3 F 0.0
10:00 225.5 F 0.0
15:00 228.8 F 0.0
20:00 230.4 F 0.0
25:00 228.7 F 0.0
30:00 229.3 F 0.0
Drying Complete

FILLING: 1:55
HEATING: 19:04
STERILIZING: 4:00
VENTING: 3:16
DRYING: 30:00
TOTAL CYCLE: 00:58:15
CYCLE COMPLETE

05/16/2023 10:27 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____