
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 001287
9:48 26/02/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 4:08
PRESSURIZING 4:08
STERILIZING 5:02
136.0 C 220kPa 5:02
136.3 C 224kPa 5:32
136.3 C 223kPa 6:02
136.4 C 223kPa 6:32
136.0 C 224kPa 7:02
136.4 C 224kPa 7:32
136.4 C 224kPa 8:02
136.4 C 223kPa 8:32
136.4 C 223kPa 8:32
VENTING 8:32
AIR DRYING 9:28
CYCLE COMPLETE 9:53

Digital Signature #
32394CD48DF60447
