
Midmark M11 - v1.0.0
Total Cycles: 905

Name: RITTER 1
Sterilizer ID: V2178198
06/16/2021 10:15 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 174.2 F 0.0
2:00 170.6 F 2.1
4:00 172.2 F 5.1
6:00 192.8 F 9.4
8:00 213.4 F 8.5
10:00 223.1 F 9.1
12:00 235.3 F 10.2
14:00 237.1 F 11.3
16:00 236.1 F 9.3
18:00 250.1 F 15.8
20:00 261.5 F 22.4
22:00 272.2 F 29.5

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 29.0
1:00 272.7 F 29.7
2:00 271.7 F 28.9
3:00 272.5 F 29.8
4:00 272.3 F 29.3

Min 271.4 F 28.9
Max 272.9 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 204.9 F 0.0
5:00 231.5 F 0.0
10:00 224.1 F 0.0
15:00 224.0 F 0.0
20:00 224.3 F 0.0
25:00 225.3 F 0.0
30:00 225.8 F 0.0
Drying Complete

FILLING: 1:22
HEATING: 22:30
STERILIZING: 4:00
VENTING: 3:17
DRYING: 30:00
TOTAL CYCLE: 01:01:09
CYCLE COMPLETE

06/16/2021 11:54 AM
ACCEPTED: X
REJECTED:
Operator: U01