
Midmark M11 - v1.0.0
Total Cycles: 2870

Name: RITTER 1
Sterilizer ID: V2178198
02/07/2023 02:20 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 196.0 F 0.0
2:00 188.5 F 4.5
4:00 197.4 F 9.2
6:00 217.3 F 7.8
8:00 228.0 F 9.1
10:00 236.0 F 9.6
12:00 241.4 F 12.1
14:00 240.3 F 10.6
16:00 254.3 F 17.7
18:00 266.4 F 25.1

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.5
1:00 272.1 F 28.8
2:00 272.0 F 29.1
3:00 271.9 F 28.6
4:00 272.0 F 29.2

Min 271.3 F 28.3
Max 272.6 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 210.0 F 0.0
5:00 232.7 F 0.0
10:00 227.8 F 0.0
15:00 228.4 F 0.0
20:00 231.4 F 0.0
25:00 229.4 F 0.0
30:00 233.3 F 0.0
Drying Complete

FILLING: 1:32
HEATING: 19:27
STERILIZING: 4:00
VENTING: 3:12
DRYING: 30:00
TOTAL CYCLE: 00:58:12
CYCLE COMPLETE

02/07/2023 03:18 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____