
Midmark M11 - v1.0.0
Total Cycles: 1238

Name: RITTER 1
Sterilizer ID: V2178198
10/07/2021 06:47 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 73.5 F 0.0
2:00 73.7 F 0.0
4:00 76.6 F 0.0
6:00 104.0 F 0.0
8:00 140.1 F 0.2
10:00 173.3 F 0.3
12:00 194.1 F 2.2
14:00 211.4 F 5.6
16:00 221.7 F 8.7
18:00 232.2 F 12.2
20:00 231.5 F 7.3
22:00 242.2 F 12.2
24:00 237.7 F 9.5
26:00 238.9 F 9.0
28:00 248.4 F 14.1
30:00 259.9 F 20.4
32:00 270.1 F 27.0

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 27.9
1:00 272.5 F 28.8
2:00 272.0 F 28.2
3:00 271.4 F 28.1
4:00 272.7 F 28.8

Min 271.2 F 27.8
Max 272.8 F 28.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 213.8 F 0.0
5:00 220.5 F 0.0
10:00 234.7 F 0.0
15:00 229.8 F 0.0
20:00 196.7 F 0.0
25:00 213.9 F 0.0
30:00 210.3 F 0.0

Drying Complete

FILLING: 0:53
HEATING: 32:58
STERILIZING: 4:00
VENTING: 3:07
DRYING: 30:00
TOTAL CYCLE: 01:10:58
CYCLE COMPLETE

10/07/2021 08:12 AM
ACCEPTED: X
REJECTED:
Operator: U02