
Midmark M11 - v1.0.0
Total Cycles: 2435

Name: RITTER 1
Sterilizer ID: V2178198
10/14/2022 03:20 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 125.3 F 0.0
2:00 129.5 F 1.5
4:00 155.9 F 4.8
6:00 179.1 F 8.9
8:00 200.7 F 13.6
10:00 220.0 F 5.9
12:00 231.9 F 10.4
14:00 234.3 F 7.7
16:00 241.6 F 12.2
18:00 239.5 F 10.3
20:00 252.5 F 16.7
22:00 263.6 F 23.4
24:00 271.7 F 28.5

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.6
1:00 272.4 F 29.2
2:00 271.5 F 28.7
3:00 272.4 F 29.2
4:00 271.4 F 28.6

Min 271.4 F 28.5
Max 272.9 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 210.9 F 0.0
5:00 222.5 F 0.0
10:00 233.8 F 0.0
15:00 234.7 F 0.0
20:00 228.1 F 0.0
25:00 225.9 F 0.0
30:00 225.3 F 0.0
Drying Complete

FILLING: 0:53
HEATING: 24:07
STERILIZING: 4:00
VENTING: 3:13
DRYING: 30:00
TOTAL CYCLE: 01:02:13
CYCLE COMPLETE

10/14/2022 04:22 PM
ACCEPTED:
REJECTED:

Reason: _____
Operator: _____