
Midmark M11 - v1.0.0
Total Cycles: 2163

Name: RITTER 1
Sterilizer ID: V2178198
07/21/2022 12:05 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 195.4 F 0.0
2:00 186.3 F 4.1
4:00 200.3 F 9.2
6:00 222.9 F 7.8
8:00 231.2 F 9.6
10:00 235.9 F 9.2
12:00 242.6 F 12.3
14:00 240.6 F 10.8
16:00 253.6 F 17.4
18:00 264.9 F 24.1

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.6
1:00 272.3 F 29.0
2:00 271.4 F 28.7
3:00 272.4 F 29.0
4:00 271.3 F 28.6

Min 271.3 F 28.4
Max 272.5 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 199.3 F 0.0
5:00 232.7 F 0.0
10:00 218.2 F 0.0
15:00 212.9 F 0.0
20:00 215.1 F 0.0
25:00 222.6 F 0.0
30:00 229.4 F 0.0

Drying Complete

FILLING: 2:04
HEATING: 19:53
STERILIZING: 4:00
VENTING: 3:16
DRYING: 30:00
TOTAL CYCLE: 00:59:13
CYCLE COMPLETE

07/21/2022 01:04 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____