
Midmark M11 - v1.0.5
Total Cycles: 2417

Name: RITTER2
Sterilizer ID: V2313228
05/25/2023 01:15 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 201.3 F 0.0
2:00 198.1 F 4.3
4:00 206.2 F 9.7
6:00 223.1 F 6.4
8:00 235.3 F 10.8
10:00 234.2 F 8.6
12:00 239.9 F 10.0
14:00 248.5 F 14.7
16:00 260.7 F 21.6
18:00 271.6 F 28.9

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.8
1:00 272.4 F 29.4
2:00 271.4 F 28.6
3:00 272.4 F 29.6
4:00 272.0 F 29.1

Min 271.2 F 28.6
Max 272.7 F 29.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 207.9 F 0.0
5:00 229.6 F 0.0
10:00 233.0 F 0.0
15:00 230.2 F 0.0
20:00 225.6 F 0.0
25:00 228.7 F 0.0
30:00 228.3 F 0.0
Drying Complete

FILLING: 2:09
HEATING: 18:37
STERILIZING: 4:00
VENTING: 3:13
DRYING: 30:00
TOTAL CYCLE: 00:57:60
CYCLE COMPLETE

05/25/2023 02:12 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____