

-----  
Midmark M11 - v1.0.0  
Total Cycles: 1767

Name: RITTER 1  
Sterilizer ID: V2178198  
03/30/2022 09:34 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 209.2 F 0.0  
2:00 200.7 F 3.3  
4:00 200.7 F 7.4  
6:00 228.6 F 13.4  
8:00 229.9 F 8.5  
10:00 239.4 F 11.1  
12:00 241.4 F 11.5  
14:00 238.9 F 9.9  
16:00 252.9 F 16.8  
18:00 264.9 F 24.0

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 28.4  
1:00 272.4 F 29.0  
2:00 271.6 F 28.2  
3:00 272.4 F 29.2  
4:00 272.2 F 28.7  
  
Min 271.4 F 28.2  
Max 272.6 F 29.2

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 213.3 F 0.0  
5:00 219.9 F 0.0  
10:00 231.5 F 0.0  
15:00 235.1 F 0.0  
20:00 236.4 F 0.0  
25:00 228.9 F 0.0  
30:00 222.9 F 0.0  
Drying Complete

FILLING: 1:23  
HEATING: 19:49  
STERILIZING: 4:00  
VENTING: 3:05  
DRYING: 30:00  
TOTAL CYCLE: 00:58:17  
CYCLE COMPLETE

03/30/2022 10:32 AM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_