

-----  
Midmark M11 - v1.0.5  
Total Cycles: 919

Name: RITTER2  
Sterilizer ID: V2313228  
03/29/2022 08:25 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 185.5 F 0.0  
2:00 181.3 F 0.3  
4:00 185.3 F 2.6  
6:00 209.7 F 7.9  
8:00 225.3 F 13.6  
10:00 233.2 F 10.0  
12:00 236.8 F 8.4  
14:00 241.3 F 10.9  
16:00 247.6 F 14.4  
18:00 260.5 F 21.7  
20:00 272.0 F 29.3

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 29.0  
1:00 272.8 F 29.7  
2:00 271.8 F 28.9  
3:00 272.3 F 29.7  
4:00 272.7 F 29.5

Min 271.5 F 28.8  
Max 272.9 F 29.8

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 207.5 F 0.0  
5:00 215.5 F 0.0  
10:00 228.3 F 0.0  
15:00 227.5 F 0.0  
20:00 227.5 F 0.0  
25:00 224.7 F 0.0  
30:00 229.8 F 0.0  
Drying Complete

FILLING: 0:59  
HEATING: 20:34  
STERILIZING: 4:00  
VENTING: 3:05  
DRYING: 30:00  
TOTAL CYCLE: 00:58:38  
CYCLE COMPLETE

03/29/2022 09:23 AM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_