
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
1.3uS / 0.8ppm
CYCLE NUMBER 001095
15:46 25/05/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:31
PRESSURIZING 3:38
STERILIZING 5:04
136.3 C 220kPa 5:04
136.3 C 225kPa 5:34
136.4 C 225kPa 6:04
136.2 C 224kPa 6:34
136.4 C 226kPa 7:04
136.4 C 225kPa 7:34
136.4 C 225kPa 8:04
136.4 C 223kPa 8:34
136.4 C 223kPa 8:34
VENTING 8:34
AIR DRYING 9:34
CYCLE COMPLETE 14:29

Digital Signature #
68EE3BBD859965F3
