
Midmark M11 - v1.0.0
Total Cycles: 1575

Name: RITTER 1
Sterilizer ID: V2178198
01/31/2022 03:45 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 155.4 F 0.0
2:00 158.7 F 3.3
4:00 184.4 F 7.3
6:00 210.0 F 12.8
8:00 220.6 F 6.5
10:00 235.7 F 11.6
12:00 233.9 F 8.5
14:00 241.4 F 10.5
16:00 247.0 F 13.5
18:00 259.4 F 20.4
20:00 270.8 F 27.7

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.2
1:00 272.5 F 28.6
2:00 271.7 F 28.4
3:00 272.4 F 28.7
4:00 271.4 F 28.1

Min 271.4 F 28.1
Max 272.9 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 208.6 F 0.0
5:00 228.8 F 0.0
10:00 231.7 F 0.0
15:00 223.7 F 0.0
20:00 224.0 F 0.0
25:00 229.8 F 0.0
30:00 232.3 F 0.0
Drying Complete

FILLING: 1:15
HEATING: 20:43
STERILIZING: 4:00
VENTING: 3:14
DRYING: 30:00
TOTAL CYCLE: 00:59:12
CYCLE COMPLETE

01/31/2022 04:44 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____