
Midmark M11 - v1.0.5
Total Cycles: 3699

Name: RITTER2
Sterilizer ID: V2313228
06/14/2024 08:24 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 71.5 F 0.0
2:00 74.0 F 0.0
4:00 83.8 F 0.6
6:00 111.9 F 2.6
8:00 148.1 F 5.5
10:00 174.0 F 8.7
12:00 190.5 F 12.8
14:00 207.4 F 4.1
16:00 217.6 F 7.4
18:00 228.8 F 11.4
20:00 229.4 F 7.4
22:00 241.9 F 12.8
24:00 240.5 F 11.1
26:00 252.8 F 17.3
28:00 263.7 F 23.8
30:00 272.0 F 29.0

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 29.0
1:00 272.9 F 29.6
2:00 271.7 F 29.1
3:00 272.8 F 29.5
4:00 271.6 F 29.0

Min 271.5 F 28.8
Max 273.1 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 199.7 F 0.0
5:00 201.6 F 0.0
10:00 201.5 F 0.0
15:00 192.9 F 0.0
20:00 192.4 F 0.0
25:00 197.4 F 0.0
30:00 203.2 F 0.0
Drying Complete

FILLING: 0:49
HEATING: 30:09
STERILIZING: 4:00
VENTING: 3:15
DRYING: 30:00
TOTAL CYCLE: 01:08:14
CYCLE COMPLETE

06/14/2024 09:32 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____