

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.4uS / 0.2ppm  
CYCLE NUMBER 000944  
13:46 18/02/2022  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:24  
PRESSURIZING 3:29  
STERILIZING 4:43  
136.0 C 221kPa 4:43  
136.4 C 233kPa 5:13  
136.3 C 231kPa 5:43  
136.1 C 230kPa 6:13  
136.3 C 232kPa 6:43  
136.4 C 229kPa 7:13  
136.4 C 233kPa 7:43  
136.3 C 229kPa 8:13  
136.3 C 229kPa 8:13  
VENTING 8:13  
AIR DRYING 9:09  
CYCLE COMPLETE 24:09

-----  
Digital Signature #  
C600DBD12F802D4C  
-----