
Midmark M11 - v1.0.0
Total Cycles: 4310

Name: RITTER 1
Sterilizer ID: V2178198
02/16/2024 12:48 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 165.9 F 0.0
2:00 167.0 F 2.8
4:00 184.1 F 6.4
6:00 204.5 F 11.3
8:00 220.4 F 4.8
10:00 231.4 F 9.3
12:00 238.8 F 10.9
14:00 241.2 F 11.3
16:00 238.3 F 9.8
18:00 251.8 F 16.1
20:00 262.7 F 22.3
22:00 272.5 F 28.9

STERILIZING
mm:ss Degrees PSI
0:00 271.0 F 28.2
1:00 272.7 F 29.1
2:00 271.6 F 28.6
3:00 272.5 F 28.7
4:00 272.6 F 29.2

Min 270.9 F 28.1
Max 272.9 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 197.7 F 0.0
5:00 236.2 F 0.0
10:00 222.3 F 0.0
15:00 230.5 F 0.0
20:00 225.6 F 0.0
25:00 218.1 F 0.0
30:00 221.0 F 0.0
Drying Complete

FILLING: 1:32
HEATING: 22:27
STERILIZING: 4:00
VENTING: 3:49
DRYING: 30:00
TOTAL CYCLE: 01:01:47
CYCLE COMPLETE

02/16/2024 01:49 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____